


クロピドグレル硫酸塩錠（日本薬局方原案記載条件）

Clopidogrel Sulfate Tablets (The draft for the Japanese Pharmacopoeia)


H130716G

(A) Standard solution*¹

(0.133 mg/mL Isopropyl benzoate, 0.1 mg/mL Clopidogrel)





	System suitability requirement	Result
Resolution (1,2)	≥4	13.1
Relative standard deviation of the peak area ratio 1 to 2	≤1.0%	0.03%



Isopropyl benzoate (I.S.)

(B) Sample solution*²

(0.133 mg/mL Isopropyl benzoate, 0.1 mg/mL Clopidogrel)



Clopidogrel sulfate

Column : YMC-Pack Pro C18 (5 μm, 12 nm)
150 X 4.0 mmI.D.

Eluent : A) buffer*³/methanol (19/1)
B) acetonitrile/methanol (19/1)
A/B (3/2)
*³ Dissolve 0.87 g of sodium 1-pentanesulfonate in 1000 mL water, adjust pH 2.5 with H₃PO₄

Flow rate : 1.1 mL/min (adjust the flow rate so that the retention time of Clopidogrel is about 8 min)

Temperature : 30°C

Detection : UV at 220 nm

Injection : 10 μL

(The draft for the Japanese Pharmacopoeia; Assay)

*¹ Standard solution was prepared from Clopidogrel sulfate supplied as a reagent for laboratory use.

*² Sample solution was prepared from Clopidogrel sulfate tablets.